Certificate of Calibration

Issued by: Megger Instruments Limited

Dover

Archcliffe Road

Kent CT17 9EN www.megger.com

Instrument Details

Model MTR105

Description Megger Baker MTR105 Serial Number 1010361102005047

Environment Details

 Temperature
 Relative Humidity
 Mains Frequency
 Mains Voltage
 Date Tested
 6-Aug-2020

 20°C ± 3°C
 50% ± 20%
 50Hz ± 0.5Hz
 230V - 2% + 10%
 Date Commissioned

Megger.

Traceability Information

Serial No Test Equipment Description Procedure Version

T3246 MTR105 Test Equipment V22

Test Title	Applied Value	Reading Value	Tolerance	Uncertainties	Pass/Fail
Voltage Ranges					
70 mV DC	70 mV	70.0 mV	1.6 mV	±1% ±1d	Pass
7 V AC	7 V	6.99 V	0.16 V	±1% ±1d	Pass
70 V DC	70 V	70 V	1.6 V	±1% ±1d	Pass
50 Hz	50 Hz	50.0 Hz	0.3 Hz	±1% ±1d	Pass
230 V DC	230 V	230 V	6 V	±1% ±1d	Pass
Continuity Range					
4 Ω	4.292 Ω	4.08 Ω	0.14 Ω	±0.5% ±1d	Pass
50 Ω	50.289 Ω	50.2 Ω	1.7 Ω	±0.5% ±1d	Pass
500 Ω	500.31 Ω	501 Ω	27 Ω	±0.5% ±1d	Pass
5 kΩ	4996 Ω	5.04 kΩ	0.27 kΩ	±0.5% ±1d	Pass
Ω S/C Current	2.292 Ω	203.755 mA	-5 mA + 15 mA	±0.5% ±1d	Pass
Insulation Ranges					
50 kΩ @ 50 V	50.05 kΩ	49 kΩ	1.2 kΩ	±0.5% ±1d	Pass
10 MΩ @ 50 V	$9.983~M~\Omega$	9.95 ΜΩ	0.4 ΜΩ	±0.5% ±1d	Pass
O/P Volts @ 50 V / 1 mA	50 V	50.86 V	50 V - 53 V	±0.5% ±1d	Pass
0.5 MΩ @ 500 V	500 MΩ	0.50 ΜΩ	0.03 MΩ	±0.5% ±1d	Pass
100 MΩ @ 500 V	100.1 MΩ	99.9 MΩ	0.04 ΜΩ	±0.5% ±1d	Pass
O/P Volts @ 500 V / 1 mA	500 V	504.06 V	500 V - 512 V	±0.5% ±1d	Pass
100 MΩ @ 1000 V	100.1 MΩ	100 MΩ	4 ΜΩ	±0.5% ±1d	Pass
1.0 MΩ @ 1000 V	1.0003 MΩ	0.99 ΜΩ	0.04 ΜΩ	±0.5% ±1d	Pass
O/P Volts @ 1000 V / 1 mA	1000 V	1007.47 V	1000 V - 1022 V	±0.5% ±1d	Pass
Capacitance Range					
700 nF	697.1 nF	696.119 nF	37 nF	±0.5% ±1d	Pass
10 uF	10.2 uF	10.074 uF	0.7 uF	±0.5% ±1d	Pass
Four Wire Resistance Range					
1 mΩ	1.1 mΩ	1.02 mΩ	0.08 mΩ	±0.05% ±1d	Pass
5 Ω	5.0003 Ω	4.999 Ω	0.01 Ω	±0.05% ±1d	Pass
Inductance Range					
100 uH @ 120 Hz	103 uH	105.973 uH	6 uH	±2% ±1d	Pass
1 H @ 1 KHz	0.8548 H	0.858 H	0.05 H	±2% ±1d	Pass
Temperature Range					
100 DegC	100 DegC	100.4 DegC	3 DegC	±2% ±1d	Pass